


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMP-1-IBM-1*

S.N.: *2102*

Date: *NOV 23 1984*



- a) Output signal amplitude:
0 TO ± 5 VOLTS.
- b) Pulse width:
0 TO 6 NSEC.
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 10 MHz.
- f) Jitter, stability:
OK.
- g) Prime power:
+24 ± 1.0 VOLT

*5.0 VOLTS / DIV
1.0 NSEC / DIV
500 KHz.*

250 mA (MAX)

