


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MP-AV-1-C-PN-05-M*

S.N.: *2104*

Date: *JAN 10 1985*



- a) Output signal amplitude:
0 TO ± 10 VOLTS
- b) Pulse width:
0 TO 4 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*10 VOLTS/DIV
1.0 NSEC/DIV
100 KHz*

*120 / 240 V
50 - 60 Hz*

AJ