


PULSE GENERATOR

PERFORMANCE CHECK


Model: AVMN-1-PS-PN

S.N.: 2107

Date: SEPT 20, 1984



(A) 1 NSEC/DIV  
5 VOLTS/DIV  
5 MHz



(B) 5 NSEC/DIV  
5 VOLTS/DIV  
50 MHz

a) Output signal amplitude:

0 TO ±5 VOLTS

b) Pulse width:

0.2 TO 5 NSEC

c) Rise time:

≤ 100 PSEC

d) Fall time:

≤ 200 PSEC

e) PRF:

0 TO 50 PMHZ

f) Jitter, stability:

OK

g) Prime power:

110/220 V,

50-60 HZ