


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM-1-C-PN-D-6A-EW-6D*

S.N.: *2109*

Date: *NOV 29 84*



- a) Output signal amplitude:
0 TO ±5 VOLTS.
- b) Pulse width:
0 TO 6 NSEC.
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

110/220V
50-60 Hz.

5 VOLTS/DIV
1.0 NSEC/DIV
5 MHz