


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVO-1

S.N.: 2110

Date: NOV 1, 1984



- a) Output signal amplitude:
+25 TO +100 AMPS
- b) Pulse width:
25 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

~ 65 AMPS/DIV
10 NSEC/DIV
1.0 KHZ

+15 ± 1.0 VOLTS
200 mA (MAX)