


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVO-1

S.N.: 2111

Date: Nov 1, 1984



- d) Output signal amplitude:
+25 TO +100 AMPS
- e) Pulse width:
25 NSEC.
- f) Rise time:
NA
- g) Fall time:
NA
- h) PRF:
0 TO 1 KHz.
- i) Jitter, stability:
OK
- j) Prime power:

~ 6.5 AMPS/DIV
10 NSEC/DIV
1.0 KHz

+15 ± 1.0 VOLT
200 mA (MAX)