


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-N*

S.N.: *2113*

Date: *JAN 3 85*



- a) Output signal amplitude:
0 TO -5V
- b) Pulse width:
0 TO 6 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

+24 ± 1.0 V

*500 mA
(MAX).*

*5 VOLTS/DIV
1.0 NSEC/DIV
5.0 MHz.*