


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVI-L1

S.N.: 2119

Date: DEC 14 84



a) Output signal amplitude:

7.5 ± 4.0 VOLTS

b) Pulse width:

1 TO 100 NSEC.

c) Rise time:

≤ 300 PSEC.

d) Fall time:

≤ 2 NSEC.

e) PRF:

0 TO 50 KHz.

f) Jitter, stability:

SEE FIG 182.

OK

g) Prime power:

± 15 ± 1.0 VOLT

75 mA (MAX)

1.0 NSEC/DIV

32 VOLTS/DIV

50 KHz.

*[Signature]*