


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-HVI-P

S.N.: 2128

Date: SEPT 19, 1984



- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width:
≤ 1 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:

100 VOLTS/DIV
1.0 NSEC/DIV
50 KHz

+15 ± 1.0 VOLTS
200 mA (MAX)