


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-P.*

S.N.: *2131*

Date: *NOV 22 84*



- a) Output signal amplitude:
0 TO +30V...
- b) Pulse width:
≤ 1 NSGZ
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

32 VOLTS/DIV
1.0 NSGZ/DIV
500 KHz.

+15 ± 1.0 VOLTS
200 mA (MAX)