


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVO-3A-P

S.N.: 2143

Date: OCT 9/84



- a) Output signal amplitude:
+25 TO +75 AMPS
- b) Pulse width:
≤ 2.0 μ SEC
- c) Rise time: (HALF POWER)
NA
- d) Fall time:
NA
- e) PRF:
0 TO 20 KHZ.
- f) Jitter, stability:
OK.
- g) Prime power:

≈ 20 AMP/DIV
2.0 μ SEC/DIV
 $R_L = 5.1 \Omega$

+15 ± 1.0 VOLT
250 mA (MAX)