


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVO-2-M-P

S.N.: 2144

Date: OCT 9/84



a) Output signal amplitude:
0 TO +10 AMPS

b) Pulse width:
0 TO 15 NSEC.

c) Rise time:
≤ 1 NSEC

d) Fall time:
≤ 2 NSEC.

e) PRF:
0 TO 20 KHZ.

f) Jitter, stability:
OK.

g) Prime power:
+15 ± 1.0 VOLT
350 mA (MAX)

≈ 7 AMPS/DIV
5.0 NSEC/DIV.
20 KHZ.