


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVN-1-C-PN

S.N.: 2149

Date: DEC 5, 1984.



5 VOLTS/DIV
1.0 NSEC/DIV
200 MHz.

- a) Output signal amplitude:
0 TO $\pm 5V$
- b) Pulse width:
0.2 TO 10 NSEC
- c) Rise time:
 ≤ 100 PSEC
- d) Fall time:
 ≤ 100 PSEC
- e) PRF:
50 TO 250 MHz
- f) Jitter, stability:
OK
- g) Prime power:

110/220V

50-60 Hz