


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PVI-P*

S.N.: *2155*

Date: *FEB 11 85*



- a) Output signal amplitude:
0 TO +40 V
- b) Pulse width:
2 TO 100 NSER
- c) Rise time:
≤ 350 PSEZ
- d) Fall time:
≤ 2 NSER
- e) PRF:
0 TO 1 MHz
(USE P14.2)
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 1.0 V
150 mA MAX.

500 PSEZ / DIV
32 VOLTS / DIV
100 KHz.
R_L = 50Ω