


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-20-T-N*

S.N.: *2157*

Date: *JAN 2 85*



- a) Output signal amplitude:
0 TO 350 V
- b) Pulse width:
0 TO 100 NSEC
- c) Rise time:
1.5 TO 10 NSEC
- d) Fall time:
3 TO 10 NSEC
- e) PRF:
0 TO 5 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

≈ 320 V / DIV
≈ 10 NSEC / DIV
1.0 KHZ

120 / 240 V
50 - 60 HZ