


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVG-1-P*
S.N.: *2163*
Date: *NOV 12 1984*



- a) Output signal amplitude:
0 TO +250 VOLTS
- b) Pulse width:
TO 50 Ω
≤ 4 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:

100 VOLTS / DIV
5 NSEC / DIV
50 KHz
R_L = 50 Ω

+15 ± 1.0 VOLT
250 mA (MAX)

- h) PROPAGATION DELAY:
≤ 28 NSEC