


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-V-1P*
S.N.: *2165*
Date: *NOV 8 1984*



- a) Output signal amplitude:
0 TO +50 V.
- b) Pulse width:
0 TO 100 nSEC
- c) Rise time:
≤ 300 pSEC.
- d) Fall time:
≤ 2 nSEC.
- e) PRF:
0 TO 20 kHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1.0 VOLT

≈ 32 VOLTS/DIV
10 nSEC/DIV
10 kHz

325 mA
(max)
A