


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVMN-1-N

S.N.: 2167

Date: DEC 13, 1984



- a) Output signal amplitude:
0 TO -5V
- b) Pulse width:
0 TO 6 NSEC.
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 200 PSEC
- e) PRF:
0 TO 50 MHz.
- f) Jitter, stability:
OK
- g) Prime power:

5 VOLTS/DIV
1.0 NSEC/DIV
5.0 MHz

+24 ± 1.0 VOLT
500 mA (MAX)