


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVMN-1-P.

S.N.: 2169

Date: DEC 13, 1984



- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
0 TO 6 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 200 PSEC
- e) PRF:
0 TO 50 MHz
- f) Jitter, stability:
OK
- g) Prime power:

5 VOLTS/DIV

1.0 NSEC/DIV

5.0 MHz

+24 ± 1.0 VOLT

500 mA (MAX)

[Signature]