


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH-5-1-C-P

S.N.: 2172

Date: JAN 24 85



- a) Output signal amplitude:  
0 TO +10 VOLTS
- b) Pulse width:  
≤ 135 PSEC
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
0 TO 1 MHz
- f) Jitter, stability:  
OK
- g) Prime power:

10 VOLTS/DIV  
100 PSEC/DIV  
100 KHz

120/240 V  
50-60 Hz