


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVC-C-NIT1*

S.N.: *2192*

Date: *JAN 24 85*



- a) Output signal amplitude:
> 20 Volts (P-P)
- b) Pulse width:
T_i 3-10 nsec
- c) Rise time:
≤ 1 nsec
- d) Fall time:
≤ 1 nsec
- e) PRF:
0 TO 500 kHz
- f) Jitter, stability:
OK
- g) Prime power:

32 VOLTS/DIV

5 nsec/DIV

100 kHz PRF

120/240 V

50-60 Hz

CAUTION

*UNIT SHIPPED
WITH VOLTAGE
SELECTOR CARD IN
THE 240 V
POSITION.*