


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM-2 P*

S.N.: *2203*

Date: *FEB 20 85*



- a) Output signal amplitude:  
*0 TO +15V*
- b) Pulse width:  
*0 TO 2 NSEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*+24 ± 1.0 VOLT*  
*600mA (MAX)*

*10 VOLTS/DIV*  
*1.0 NSEC/DIV*  
*5 MHz*