


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APP A-14V3-P*

S.N.: *2204*

Date: *FEB 20 85*



a) Output signal amplitude:

0 TO +40 VOLTS

b) Pulse width:

0 TO 2 NSEC

c) Rise time:

≤ 150 NSEC

d) Fall time:

≤ 250 NSEC

e) PRF:

0 TO 1000 Hz

f) Jitter, stability:

OK

g) Prime power:

+15 ± 1.0 VOLT

200 mA (MAX)

32 VOLTS/DIV

1.0 NSEC/DIV

5.00 KHz