


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVP-AV-1403-C-N

S.N.: 2213

Date: DEC 13 84



) Output signal amplitude:

0 to 40V

) Pulse width:

0 to 2 nsec

) Rise time:

≤ 150 psec

) Fall time:

≤ 250 psec

e) PRF:

0 to 1 MHz

f) Jitter, stability:

OK

g) Prime power:

120/240V

50-60 Hz

32 VOLTS/DIV
1.0 nsec/DIV
100 kHz