


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVD-1-P*
S.N.: *2220*
Date: *FEB 20 85*



- a) Output signal amplitude:
25 To 100 μ amps
- b) Pulse width:
20 \pm 5 nsec
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 To 1 kHz
- f) Jitter, stability:
OK
- g) Prime power:

*\approx 60 Amps/div
10 nsec/div
500 Hz.*

*+15 \pm 1.0 VOLT
200 μ A (MAX)
*[Signature]**