

PULSE GENERATOR
PERFORMANCE CHECK

Model:


AVO-1-P

S.N.:

2221

Date:

FEB 20 85



a) Output signal amplitude:

25 to 100 AMPs

b) Pulse width:

20 ± 5 nsec

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 to 1 kHz

f) Jitter, stability:

OK

g) Prime power:

+15 ± 1.0. VOLT

200 mA (MAX)

60 AMPs/DIV

10 nsec/DIV

500 Hz.