


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVR-1-P

S.N.: 2234

Date: JAN 7 85



- a) Output signal amplitude:
0 TO +200 V
- b) Pulse width:
0.1 TO 5.0 μ SEC.
- c) Rise time:
 ≤ 10 NS/SEC
- d) Fall time:
 ≤ 10 NS/SEC
- e) PRF:
0 TO 1 KHZ.
- f) Jitter, stability:
OK
- g) Prime power:

50 VOLTS/DIV
1.0 μ SEC/DIV
1.0 KHZ.

+24 ± 1.0 VOLT
50 mA (MAX)