

PULSE GENERATOR

PERFORMANCE CHECK

Model:

ANM-1-C-PN-A-ED

S.N.:

2242

Date:

MAR 11 85

a) Output signal amplitude:

0 TO  $\pm 5$  VOLTS

b) Pulse width:

0 TO 6 NSEC.

c) Rise time:

$\leq 100$  PSEC

d) Fall time:

$\leq 135$  PSEC.

e) PRF:

0 TO 2.5 MHz


f) Jitter, stability:

OK

g) Prime power:

120/240 VOLTS

50-60 Hz.



5 VOLTS/DIV.

1.0 NSEC/DIV

5.0 MHz