


PULSE GENERATOR
PERFORMANCE CHECK


Model: *ANMR-2-C-PN-TRF*

S.N.: *2254*

Date: *JAN 22 85*



*TR, TF HIGH
5 μ SEC/DIV
10 VOLTS/DIV
5 MHz*



*TR, TF LOW
.5 μ SEC/DIV
10 VOLTS/DIV
5 MHz*

- a) Output signal amplitude:
0 TO \pm 20 VOLTS
- b) Pulse width:
20 TO 200 NSEC.
- c) Rise time:
 \leq 0.3 NSEC OR 3 NSEC
- d) Fall time:
 \leq 0.3 NSEC OR 3 NSEC.
- e) PRF:
0 TO 10 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*120/240 V
50-60 Hz.*