


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVO-5.B-P-PS

S.N.: 2267

Date: APRIL 10 85



~ 13 AMPS/DIV.  
100 NSER/DIV  
500 Hz.

( $R_L = 0.7 \Omega$ ,  
10 V/DIV)

- a) Output signal amplitude:  
0 TO + 48 AMPS
- b) Pulse width:  
25 TO 500 NSER.
- c) Rise time:  
~ 10 NSER
- d) Fall time:  
~ 10 NSER
- e) PRF:  
0 TO 1 KHz
- f) Jitter, stability:  
OK
- g) Prime power:  
120/240 V,  
50-60 Hz.