


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVK-A-C-PN-M*

S.N.: *2271*

Date: *MAR 12 85*



a) Output signal amplitude:

0 TO ± 15 VOLTS

b) Pulse width:

2 TO 100 NSECS

c) Rise time:

SEE FIGS 1,2
≤ 100 PSEC

d) Fall time:

≤ 2 NSEC

e) PRF:

0 TO 1 MHz

f) Jitter, stability:

SEE FIGS 1,2
OK

g) Prime power:

120/240 V

50-60 Hz

AK

10 VOLTS/DIV

1.0 NSEC/DIV

100 KHz