


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-1-P*
S.N.: *2272*
Date: *JAN 8 85*



- a) Output signal amplitude: *25 TO 100 AMPS*
- b) Pulse width: *25 NSEC*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 1 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *OK*

≈ 60 AMPS/DIV
10 NSEC/DIV
1 KHz

+15 ± 1.0 VOLT.
200 mA (max)