


PULSE GENERATOR

PERFORMANCE CHECK


Model: *ACMMN-1-C-PN-M*

S.N.: *2273*

Date: *MAR 6 85*



*50 MHz  
5 VOLTS/DIV  
5 NSER/DIV*



*15 MHz  
5 VOLTS/DIV  
1.0 NSER/DIV*

- a) Output signal amplitude:  
*0 TO ± 5 VOLTS*
- b) Pulse width:  
*0 TO 6 NSER*
- c) Rise time:  
*≤ 100 PSER*
- d) Fall time:  
*≤ 200 PSER*
- e) PRF:  
*0 TO 50 MHz*
- f) Jitter, stability:  
*OK.*
- g) Prime power:

*120/240 V  
50-60 Hz*

