


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *APD-M-1-C-DN*

S.N.: *2278*

Date: *MAY 18 85*



- ) Output signal amplitude:  
*0 TO ± 10 VOLTS*
- ) Pulse width:  
*0 TO 4 NSEC.*
- ) Rise time:  
*≤ 100 PSEC.*
- ) Fall time:  
*≤ 135 PSEC.*
- ) PRF:  
*0 TO 100 kHz*
- ) Jitter, stability:  
*OK*
- ) Prime power:

*10 VOLTS/DIV  
1.0 NSEC/DIV  
100 kHz*

*120/240 V  
50-60 Hz.*  
*[Signature]*