


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *APR-1-PW-C-EA-EW-OS-PN*

S.N.: *2279*

Date: *MAR 27 85*



P OUT.  
*50 VOLTS/DIV*  
*1.0 USEC/DIV*  
*500 Hz*



N OUT.  
*50 VOLTS/DIV*  
*1.0 USEC/DIV*  
*500 Hz*

- a) Output signal amplitude:  
*0 TO ± 200 VOLTS.*
- b) Pulse width:  
*0.1 TO 5.0 USEC*
- c) Rise time:  
*≤ 10 NSEC*
- d) Fall time:  
*≤ 10 NSEC*
- e) PRF:  
*0 TO 1 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*120/240 V*  
*50-60 Hz*

*[Signature]*