


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-C-PN-D-M*

S.N.: *2280*

Date: *MAR 18 85*



- a) Output signal amplitude:
0 TO ±5V OUTS.
- b) Pulse width:
0 TO 6 NSER
- c) Rise time:
≤ 100 PSEZ
- d) Fall time:
≤ 135 PSEZ
- e) PRF:
0 TO 25 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V

*5 VOUTS / DIV
100 NSER / DIV
5.0 KHz.*

50-60 Hz.

[Signature]