


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-17*

S.N.: *2298*

Date: *FEB 4 85*



a) Output signal amplitude:
0 TO +30 VOLTS

b) Pulse width:
≤ 1 NSEC.

c) Rise time:
NA

d) Fall time:
NA

e) PRF:
0 TO 100 KHz

f) Jitter, stability:
OK

g) Prime power:
+15 ± 1.0 VOLTS
150 mA (MAX)

16 VOLTS/DIV
100 NSEC/DIV
100 KHz