


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AP-AV-1-C-PN-05-M*

S.N.: *2300*

Date: *NOV 8 83*



40 dB ATTEN

10 V/div

PRF = 1.0 MHz

- a) Output signal amplitude:
0 to ± 100 V
- b) Pulse width:
0.2 to 4.0 nsec
- c) Rise time:
≤ 100 psec
- d) Fall time:
≤ 135 psec
- e) PRF:
0 to 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
110/220 V
50-60 Hz


VM : OK
OS : OK

PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-AV-1-C-PN-05-M*

S.N.: *2300*

Date: *MAR 18 85*



- a) Output signal amplitude:
0 TO ± 10 VOLTS
- b) Pulse width:
0 TO 4 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*10 VOLTS/DIV
1.0 NSEC/DIV
100 kHz*

*120/240 V
50-60 Hz*