


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AV-HV2-C-PN-05*

S.N.: *2308*

Date: *APRIL 22 85*



- a) Output signal amplitude:
0 TO ± 20 VOLTS
- b) Pulse width:
0 TO 2 NSEC
- c) Rise time:
≤ 100 PSEC.
- d) Fall time:
≤ 200 PSEC
- e) PRF:
0 TO 100 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*10 VOLTS/DIV
1.0 NSEC/DIV
100 KHz.*

*120/240 V
50 - 60 Hz.*