


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH-MV1-P

S.N.: 2311

Date: APRIL 22 85



- a) Output signal amplitude: 0 TO +100 VOLTS
- b) Pulse width:  $\leq 1 \mu\text{SEC}$
- c) Rise time: NA
- d) Fall time: NA
- e) PRF: 0 TO 100 KHZ
- f) Jitter, stability: OK
- g) Prime power: OK

100 VOLTS/DIV  
1.0  $\mu\text{SEC}$ /DIV  
50 KHZ.

+15  $\pm$  1.0 V  
150 mA  
MAX  
JAB