


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMAN-1-P*

S.N.: *2316*

Date: *FEB 13 85*



- a) Output signal amplitude:
0 TO 45 VOLTS
- b) Pulse width:
0 TO 6 NSOR
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 200 PSEC
- e) PRF:
0 TO 50 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*5 VOLTS/DIV
100 NSOR/DIV
40 MHz.*

*+24 ± 1.0 V
450 mA (MAX)*