


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMIN-1-P*

S.N.: *2317*

Date: *FEB 13 85*



- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
0 TO 6 USEC.
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 200 PSEC
- e) PRF:
0 TO 50 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+24 ±1.0 VOLTS
450 mA (MAX)

5 VOLTS / DIV
1.0 USEC / DIV
40 MHz