


PULSE GENERATOR
PERFORMANCE CHECK

Model: *VAVMN-1-N*

S.N.: *2318*

Date: *FEB 13 85*



- a) Output signal amplitude:
0 TO - 5 VOLTS
- b) Pulse width:
0 TO 6 USEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 200 PSEC
- e) PRF:
0 TO 50 MHz
- f) Jitter, stability:
OK
- g) Prime power:
OK

+24 ± 1.0 VOLTS

*450 mA
(MAX)*

5 VOLTS/DIV

1.0 USEC/DIV

40 MHz