

PULSE GENERATOR

PERFORMANCE CHECK

Model:


AP-V-2 -P-TT

S.N.:

2320

Date:

FEB 26 85



- d) Output signal amplitude:  
+10V
- e) Pulse width:  
2 TO 50 NS
- f) Rise time:  
≤ 100 PSEC
- g) Fall time:  
≤ 135 PSEC
- h) PRF:  
0 TO 50 KHZ
- i) Jitter, stability:  
OK
- j) Prime power:

10 VOLTS/DIV

5 NSEC/DIV

50 KHZ

+15 ±1.0V

300 mA MAX