


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVR-3-PW-C-P

S.N.: 2327

Date: MAY 16 '85



- a) Output signal amplitude:  
0 TO +200 V.
- b) Pulse width:  
0.1 TO 100  $\mu$ SEC
- c) Rise time:  
 $\leq$  10 NSEC
- d) Fall time:  
 $\leq$  10 NSEC
- e) PRF:  
MIN PW : 0 TO 1 KHz  
MAX PW : 0 TO 50 Hz.
- f) Jitter, stability:  
OK.
- g) Prime power:

50 VOLTS/DIV  
10  $\mu$ SEC/DIV  
 $R_L = 50 \Omega$   
200 Hz.

120/240 V  
50 - 60 Hz.