

PULSE GENERATOR  
PERFORMANCE CHECK

Model:

PVL-2-CT

S.N.:

2330

Date:

APRIL 8 85

a) Output signal amplitude:

0 TO + 350 VOLTS

b) Pulse width:

0 TO 100 NSEC

c) Rise time:

$\leq 1.5$  NSEC

d) Fall time:

$\leq 3$  NSEC

e) PRF:

0 TO 5 KHZ


f) Jitter, stability:

OK

g) Prime power:

110/220 V

50-60 HZ



$\approx 320$  VOLTS/DIV  
10 NSEC/DIV  
2 KHZ