


PULSE GENERATOR  
PERFORMANCE CHECK

Model: PVL-2-C-N

S.N.: 2331

Date: APRIL 8 85



- a) Output signal amplitude:  
0 to 350 V
- b) Pulse width:  
0 to 100 nsec
- c) Rise time:  
≤ 1.5 nsec
- d) Fall time:  
≤ 3 nsec
- e) PRF:  
0 to 5 kHz
- f) Jitter, stability:  
OK
- g) Prime power:  
120/240 V

≈ 320 VOLTS/DIV  
10 nsec/DIV  
2 kHz

50 - 60 MHz