


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVP-100-1-C-PN-05-M*

S.N.: *2334*

Date: *PMAR 20 85*



- a) Output signal amplitude:  
*0 TO ± 10 VOLTS*
- b) Pulse width:  
*0 TO 4 USEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 1 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*120 / 240 V*  
*50-60 Hz*

*10 VOLTS/DIV*  
*1.0 USEC/DIV*  
*100 KHz*