


PULSE GENERATOR
PERFORMANCE CHECK


Model: AUMN-1-C-10

S.N.: 2337

Date: APRIL 23 85



Ⓐ 5 VOLTS/DIV
10 NSEC/DIV
PRF = 40 MHz



Ⓑ 5 VOLTS/DIV
10 NSEC/DIV
PRF = 40 MHz

- a) Output signal amplitude:
0 TO +5V
- b) Pulse width:
0 TO 6 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 200 PSEC
- e) PRF:
0 TO 50 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V
50-60 Hz