


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *AVMR-2-C-TRF-PN-M.*

S.N.: *2343*

Date: *JUNE 6 '85*



*A) 10 VOLTS/DIV  
10 NSEC/DIV  
1 MHz.  
TRF HIGH*



*B) AS A) BUT  
TRF LOW.*

- a) Output signal amplitude:  
*0 TO ± 20 VOLTS*
- b) Pulse width:  
*20 TO 200 NSEC*
- c) Rise time:  
*≤ 0.3 OR 3 NSEC*
- d) Fall time:  
*≤ 0.3 OR 3 NSEC*
- e) PRF:  
*0 TO 10 MHz*
- f) Jitter, stability: *0 TO 5 MHz IN TEST TRF (MODE)*  
*OK*
- g) Prime power:

*120/240V  
50-60 Hz*